

	<h1>PRODUCT SPECIFICATION</h1>	Doc No	
		PS 1054	
		Rev	Page
		0	2 of 2

Product : Non-Sterile, Ambidextrous, Texture Surface, Powder Free
Natural Rubber Latex Examination Gloves

Color : Natural

Product Code : OLPYT-NA-6.0 (ER36001CA-NA)

Reference Code : PS60Y

Grade : AQL 1.5

Inspection Method : GB 10213 (Current Version)

Quality Assurance : Manufactured under ISO 13485, CAN/CSA ISO 13485, ISO 9001 and US FDA QSR Quality Management System

No	Test	Particular	Specification	In-house Inspection Level / AQL	
1	Dimension (mm)	Width	Extra Small	75 ± 5	S2/AQL 4.0
			Small	85 ± 5	
			Medium	95 ± 5	
			Large	106 ± 5	
			Extra Large	116 ± 5	
		Length	Min. 240		
		Thickness (Single Wall)	Cuff (25 ± 5 from bead)	Min. 0.08	
Palm (center of palm)	Min. 0.11				
Finger (13 ± 3 from tip)	Min. 0.12				
2	Physical Properties	Unaged	Force at Break (N)	Min. 7.0	S2/AQL 4.0
			Elongation at Break (%)	Min. 650	
		Aged	Force at Break (N)	Min. 6.0	
			Elongation at Break (%)	Min. 500	
3	Freedom from Holes			GI/AQL 1.5	
4	Visual Defects (Major)	Dirt ≥1mm ² , Sticky Pleat ≥5mm, Visible Hole (Non-Inflated), Discoloration, Mix Surface, Incomplete Beading, Wrong Orientation, Deformed Gloves		GI/AQL 2.5	
5	Visual Defects (Minor)	Lump <2mm ² , Dirt <1mm ² , Sticky Pleat <5mm, Poor Beading		GI/AQL 4.0	
6	Powder Residue	Max 2.0 mg/glove		n = 5	

Distribution or copying of this document is prohibited without the prior permission of KISB